



**Reply to quarries of Pre Bid meeting of "Online GHG Monitoring System"
Held on 05-03-2018 at 1100 Hrs**


The queries received from the following two firms.

1. Aaxis Nano Technologies Pvt. Ltd.
2. Info Electronics System India Pvt. Ltd.

S. No.	Quarries	Clarification
Aaxis Nano Technologies Pvt. Ltd.		
1.	Concentration of H ₂ O from 30 ppm to less than equal to 50 ppm (in 4.2 of RFP).	As per RFP
2.	Size of dust filter from 10 micron to 7 micron. (In 4.3.2 of RFP).	Filter range 7 Micron to 10 Micron is acceptable.
Info Electronics System India Pvt. Ltd.		
3.	Some spare is not required on Piccaro analyzer as mention in RFP	It will remain same as earlier in case of government procurements from abroad.
4.	Does IMD have exemption from Custom Duty?	IMD will issue custom duty exception certificate. If it is not agreed by custom department, actual reimbursement will be made of original receipt.
5.	Delivery in two Consignments.	Agreed
6.	Rack is required for rack Mount	Rack should be quoted.
7.	Why spares is required when equipment in under warranty and CAMC. Whether vender will be allowed to use spares.	Vender has to quote for spares mentioned in tender. Vender will be allowed to use spares, however supplier firm has to replenish the spare used.


(M.R Durai)
Chairman


(Sanjay Bist)
Member Secretary


(N. Puviarasan)
Member